

## § 30.25

Register on March 23, 1981. This publication may be inspected at the Office of Federal Register, 800 North Capitol Street, NW., suite 700, Washington, DC, and is available from the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pennsylvania 19103.

(Sec. 201. Pub. L. 85-859, 72 Stat. 1358, as amended (26 U.S.C. 5204); 80 Stat. 383, as amended (5 U.S.C. 552(a)))

[T.D. ATF-198, 50 FR 8535, Mar. 1, 1985, as amended by T.D. ATF-381, 61 FR 37004, July 16, 1996]

### § 30.25 Use of precision specific gravity hydrometers.

The provisions of § 30.23 respecting the care, handling, and use of precision instruments shall be followed with respect to the care, handling, and use of precision grade specific gravity hydrometers. Specific gravity hydrometers shall be read to the nearest subdivision. Because of temperature density relationships and the selection of the standardization temperature of 60 °/ 60 °F., the specific gravity readings will be greater at temperatures below 60 degrees Fahrenheit and less at temperatures above 60 degrees Fahrenheit. Hence, correction of the specific gravity readings will be made for temperature other than 60 degrees Fahrenheit. Such correction may be ascertained by dividing the specific gravity hydrometer reading by the applicable correction factor in Table 7.

*Example:* The specific gravity hydrometer reading is 1.1525, the thermometer reading is 68 degrees Fahrenheit, and the true proof of the spirits is 115 degrees. The correct specific gravity reading will be ascertained as follows:

(a) From Table 7, the correction factor for 115° proof at 68 °F. is 0.996.

(b) 1.1525 divided by 0.996=1.1571, the corrected specific gravity.

(Sec. 201, Pub. L. 85-859, 72 Stat. 1358, as amended (26 U.S.C. 5204))

## Subpart D—Gauging Procedures

### § 30.31 Determination of proof.

(a) *General.* The proof of spirits shall be determined to the nearest tenth degree which shall be the proof used in determining the proof gallons.

## 27 CFR Ch. I (4–1–99 Edition)

(b) *Solids content not more than 600 milligrams.* Except as otherwise authorized by the Director, the proof of spirits containing not more than 600 milligrams of solids per 100 milliliters of spirits shall be determined by the use of a hydrometer and thermometer in accordance with the provisions of § 30.23 except that if such spirits contain solids in excess of 400 milligrams but not in excess of 600 milligrams per 100 milliliters at gauge proof, there shall be added to the proof so determined the obscuration determined as prescribed in § 30.32.

(c) *Solids content over 600 milligrams.* If such spirits contain solids in excess of 600 milligrams per 100 milliliters at gauge proof, the proof shall be determined on the basis of true proof determined as follows:

(1) By the use of a hydrometer and a thermometer after the spirits have been distilled in a small laboratory still and restored to the original volume and temperature by the addition of pure water to the distillate; or

(2) By a recognized laboratory method which is equal or superior in accuracy to the distillation method.

(d) *Initial proof.* Except when the proof of spirits is used in making the guage prescribed in 27 CFR 19.383 or in making a gauge for determination of tax, the initial determination of proof made on the bonded premises of a distilled spirits plant for such spirits may be used whenever a subsequent gauge is required to be made at that same plant provided that no material has been added to change the proof of the spirits.

(Sec. 201, Pub. L. 85-859, 72 Stat. 1358, as amended, 1362, as amended (26 U.S.C. 5204, 5211))

### § 30.32 Determination of proof obscuration.

(a) *General.* Proof obscuration of spirits containing more than 400 but not more than 600 milligrams of solids per 100 milliliters shall be determined by one of the following methods. The evaporation method may be used only for spirits in the range of 80–100 degrees at gauge proof.

(b) *Evaporation method.* Evaporate the water and alcohol from a carefully